



ENV AUTOMATION PTE LTD

60 Paya Lebar Road, #07-54 Paya Lebar Square, Singapore 409051  
Email: envautomation.sg@gmail.com | Website: www.autosiltstopper.com

CERTIFICATE NO  
ENV2026009

CALIBRATION CERTIFICATE

This certificate confirms that the instrument listed below has been inspected and calibrated according to internal inspection requirements and applicable calibration practice.

CLIENT INFORMATION

Samsung C&T Corporation

PROJECT NAME

CR112-ECM 4

SENSOR TYPE

TSS Sensor

SENSOR NUMBER

864650056138461

DATE OF CALIBRATION

11/09/2025

NEXT CALIBRATION

11/09/2026 VALID

| Test Items                       | Test Requirements                           | Test Result            |
|----------------------------------|---------------------------------------------|------------------------|
| Visual inspection test           | No obvious scratches, damage, ect           | <span>Qualified</span> |
| Does the screw have a leak lock  | There is no                                 | <span>Qualified</span> |
| There is no sundry in the sensor | There is no                                 | <span>Qualified</span> |
| There is no sundry in the sensor | Soak for more than 12 hours without water   | <span>Qualified</span> |
| There is no sundry in the sensor | Connect the transmitter OK                  | <span>Qualified</span> |
| There is no sundry in the sensor | High temperature test for 3 hours the above | <span>Qualified</span> |
| There is no sundry in the sensor | Oscillate 250 rpm for more than 2 hours     | <span>Qualified</span> |

| Functional Test                                                                        | Standard Solution |       | Actual Measurement |       |
|----------------------------------------------------------------------------------------|-------------------|-------|--------------------|-------|
| Calibration test<br>(mg / L : 10 % 10 mg /L)<br>(Standard)                             | Solution 1        | 0.2   | 0.53               | 0.33  |
|                                                                                        | Solution 2        | 207   | 204.97             | 1.0 % |
|                                                                                        | Solution 3        | 5070  | 5012.49            | 1.1 % |
|                                                                                        | Solution 4        | 12050 | 12237.36           | 1.6 % |
| Note : Before testing must refer to the customer requirement and inspection standards. |                   |       |                    |       |

AUTHORIZED BY



Checked by Signature

OFFICIAL COMPANY STAMP

